

[Open Peer Review on Qeios](#)

Auto Train Brain mobile app for promoting dyslexia biomarker detection in children at home or at school: Feasibility, Acceptability, Economic impact, Pilot Study and Survey Results

Gunet Eroglu, Mertali Köprülü¹, Berdan Karabacak²

¹ Isik University

² Neuro Brain Academy

Funding: No specific funding was received for this work.

Potential competing interests: No potential competing interests to declare.

Abstract

The use of mobile apps in diagnosing and improving health conditions has increased in recent years. If dyslexia is diagnosed at 7 years old, dyslexic children's reading and spelling performance get better. Therefore, dyslexia diagnosis at an early age is very important but mostly delayed in Turkey to give a chance to overcome any maturation delays. We need a biomarker app for dyslexia because it potentially helps to prevent severe consequences through an early diagnosis that helps provide early intervention. In the literature, there is no clinically tested mobile app for dyslexia biomarker detection that can objectively monitor the child's situation at home or at school. No other research assessed the feasibility and acceptability of using this kind of mobile app to detect dyslexia with a survey on dyslexic families. Here we present a novel neurofeedback mobile app that has an embedded dyslexia biomarker based on Z-scored QEEG data that can be used at home or school that has accomplished a high accuracy rate in diagnosing dyslexia. The mobile app can be used at home by parents or teachers at school. We have collected data from 207 children (96 of them have dyslexia, 111 of them are typically developing) between 7-10 years old for 60 sessions during their neurofeedback sessions. The data consists of the eyes-open resting state 2-minute QEEG data from 14-channels. Using the ANN machine learning method, dyslexic/normal classification has been done with a high accuracy rate (98.8%). ANN yields the highest accuracy results with standardized QEEG data in the literature. Auto Train Brain is a novel neurofeedback mobile app that has dyslexia biomarker detection software embedded into it. A survey is created to assess the mobile app's dyslexia biomarker detection module's feasibility, acceptability, and economic impact. The results have shown that the biomarker detection app module is found feasible and acceptable by families, however, it is found expensive to use at home.

Günet Eroğlu^a, Mertali Köprülü^b, Berdan Karabacak^c

^a Computer Engineering Dept, Engineering and Nature Faculty, Bahçeşehir University, Istanbul, Turkey

^b Computer Engineering Dept, Engineering and Nature Faculty, Işık University, Istanbul, Turkey

^c Research Department, Neuro Brain Academy, Kocaeli, Turkey

Keywords: Dyslexia detection, developmental dyslexia biomarkers, QEEG Dataset, Artificial-Neural Network, Auto Train Brain.

1. Introduction

Dyslexia is a type of learning disability that has neurological underpinnings. Dyslexia is characterized by marked hypoactivation within the reading network, disrupted functional connectivity, and differences in structural connectivity in certain fiber tracts (Kuhl et al., 2020). It's characterized by problems with accurate and fluent word recognition, as well as poor spelling skills (Snowling et al., 2012). Although IQ is normal or above normal, dyslexic people find it difficult to read. Other dyslexia-related implications include difficulties with reading comprehension and a lack of reading experience, both of which can hinder the development of vocabulary and knowledge (Lyon, Shaywitz, & B.A. Shaywitz, 2003). Because of weak connections, there is a "disconnection hypothesis" for dyslexia (Paulesu et al., 1996).

Reading is an acquired ability for humans (Dehaene, 2005). A child's visual and auditory abilities should be developed as time goes by and the left hemispheric lateralization should be completed by the age of 7 before the child starts school and learn how to read (Koenig et al., 1990). Before the diagnosis of dyslexia, additional learning support should be provided to the child and the nutritional deficiencies should be eliminated (Handler et al., 2011). The diagnosis of dyslexia is performed by psychiatrists in Turkey; they use psychometric tests, and the first diagnosis may start 2 years after the child starts school. During this period, as the child can not properly read and write, his self-esteem would go down and his academic life starts to be affected by the late intervention. With the help of computers and machine learning, this period would be shortened, and the diagnosis of dyslexia can be done at 7 years old, and the remedial interventions may be taken much beforehand. Therefore, some children may overcome dyslexia even before they start reading.

In the literature, machine learning methods have been utilized to diagnose dyslexia. These algorithms use psychometric test results, fMRI scans, EEG scans, PET scans, MEG scans, eye tracking information, face images, handwritten texts, and mobile-based games. Psychometric tests take 1.5- 2 hours, and the child should be taken to a psychiatrist. Wechsler Abbreviated Scale of Intelligence (WASI, II, III&IV) (Wechsler, 1999), Wide Range Achievement Test: Revision R,3&4 (WRAT-R,3&4) (Jastak & Wilkinson, 1984), Woodcock & Johnson III (WJ-III) (Schrack, 2011), Comprehensive Test of Phonological Processing (CTOPP) (Wagner et al.,1999), and Peabody Picture Vocabulary Test: Third Edition (PPVT-III) (Dunn & Dunn, 1965), Test of integrated language and literacy skills (TILLS test) (Nelson et al., 2016), and WISC-R tests (Kaufman, 1994) are used as psychometric tests. Although fMRI scans and eye tracking data yield high-accuracy results compared with other methods, they are expensive solutions. It is hard to use MRI scans and eye tracking methods to collect data from a 7-year-old child as they require expensive equipment. Collecting face images from children brings privacy and security issues for the child and their families. The algorithms which use EEG scans, although it is a cheaper way to collect data, do not yield a high accuracy result to go to the market in the literature.

Artificial neural network (ANN) is a computational model that consists of several processing elements that receive

inputs and deliver outputs based on their predefined activation functions. An artificial neural network (ANN) is a non-parametric machine learning which does not assume anything about the data set, it is focused on finding the discriminant function to diagnose dyslexia. The K-means algorithm identifies the k number of centroids, and then allocates every data point to the nearest cluster while keeping the centroids as small as possible. The 'means' in the K-means refers to averaging the data; that is, finding the centroid. Support Vector Machine(SVM) is a supervised machine learning algorithm used for both classification and regression. Though we say regression problems as well it's best suited for classification. The objective of the SVM algorithm is to find a hyperplane in an N-dimensional space that distinctly classifies the data points. On EEG scan data sets, Al-Barhamtoshy and Motaweh (2017) used K-means, ANN, and fuzzy logic classifiers to reach an accuracy of 89.6%, 89.6%, and 85.7 percent, respectively. The three models were shown to have an overall accuracy of 81.1%, 62% precision, 100% recall, and 76.6% F-score. Karim et al. (2013) employed a Multilayer Perceptron to detect dyslexia signs by collecting brain waves in the resting state. With eyes closed, the accuracy was 85%, and with eyes open, it was 86%. Using machine learning, Frid and Breznitz (2018) investigated the variations in ERP signals between dyslexic and proficient readers. They used SVM, ANN, and PCA to get a 78% accuracy. In the literature, the best accuracy scores reached with ANN on fMRI scans and DTI scans are 94.8% (Chimeno et al.,2014). The best accuracy results with eye tracker data were 95.6% with SVM- RFE (Benfatto et al., 2016). The best accuracy result for dyslexia detection with handwriting data was 77.6% (Spoon et al.,2019). The best accuracy results with psychometric test scores were 99% (Khan et al., 2018).

Auto Train Brain is a neurofeedback mobile app with a headset that improves the cognitive abilities of dyslexia. The efficacy was proven with a clinical trial beforehand a clinical trial and it was shown that using Auto Train Brain improves reading comprehension and reading speed (Eroğlu et al., 2020). This study aims to claim accurately diagnose dyslexia using z-scored QEEG recordings from 14-channel band power data (greater than 95 percent) of both dyslexic and healthy children. Study tests the feasibility, acceptability, and economic impact of this app with a survey on dyslexic families. This is the first research that investigated the use of dyslexia detection software in dyslexic families.

2. Methods and materials

2.1. Participants

96 children with pure dyslexia ($M_{age} = 8.85$, $SD = 1.56$; 76 males, 20 females; the ethnic group is white) who used Auto Train Brain at home on a regular basis for neurofeedback formed our experiment group. There were 111 children in the control group who were developing typically ($M_{age} = 8.80$, $SD = 1.60$; 80 males, 31 females; the ethnic group is white). The children in the experimental group were diagnosed with dyslexia by psychiatric professionals, who then recommended that their families use Auto Train Brain at home. The TILLS tests were used by psychologists and psychiatrists to examine whether the individuals met the DSM-V dyslexia criteria. The children chosen to participate in the experiment were chosen at random. The participant's primary goal in the retrospective study is to use Auto Train Brain software as a neurofeedback device at home. Before each neurofeedback session, individuals were instructed to perform

a 2-minute resting state EEG measurement for the purpose of data collection. Before going to school in the morning, the participants used Auto Train Brain.

To evaluate the socioeconomic situation of the children, a survey of their parents was conducted. The poll includes inquiries about occupation, education (primary school, high school, and university), and income (basic income 6,000 TL, mid-level income 6,000 TL to 20,000 TL, high income >20,000 TL) (staff, blue-collar workers, white-collar workers). The study's inclusion criteria were that the participants were not on medication and did not have any other comorbid disorders than dyslexia and they came from middle-SES families. They lived in many cities around Turkey.

2-minute resting-state eyes-open theta, alpha, beta1, beta2, and gamma-band powers were measured before each neurofeedback session (60 sessions per subject on average). This study uses a small sample size and multiple measures (12,420 sessions and 207 participants). The typically developing children's 2-minute resting-state eyes open QEEG data are gathered using Auto Train Brain, and the data for the experimental and control groups are balanced to have an equal number of instances in each group.

2.2. Electroencephalography (EEG) recording

Throughout the experiments, EMOTIV's EPOC+ and EPOC-X headsets were used. The internal sampling rate of the headgear was 2048 samples per second for each channel. Prior to downsampling, the EEG data to 128 samples per second for each channel, main artifacts, and alias frequencies were removed. Each of the 14 EEG channels received 128 samples per second, along with two additional channels used as controls. Before the training, the EMOTIV EPOC+ headset was calibrated on the participants' scalps using EMOTIV APP mobile applications to make sure each electrode could send high-quality EEG data. For all analyses in this work, theta (4–8 Hz), alpha (8–12 Hz), beta-1 (12–16 Hz), beta-2 (16–25 Hz), and gamma (25–45 Hz) bands of the 14-channel EEG data were recorded throughout the tests. Data from the delta (0–4 Hz) band was absent from the EMOTIV headset interfaces. It has been demonstrated that EMOTIV EPOC+ provides research-grade QEEG data (Badcock et al, 2013). In the dataset, there are 70 features (frequency band data; theta, alpha, beta1, beta2, gamma from 14 channels) that indicate whether a patient is dyslexic or not. The information is gathered using electrodes on the Auto Brain Train device (AF3, F3, F7, FC5, T7, P7, O1, O2, P8, T8, FC6, F8, F4, AF4). The labels for the characteristics reflect where the electrodes are placed. Theta, alpha, and beta are all present in the product's readings which store the neuron activity information in the specified location.

2.3. Neurofeedback treatment protocol and multi-sensory learning method

A smartphone app called Auto Train Brain uses neurofeedback and multimodal learning methods. It works with the EMOTIV EPOC+ and EPOC-X headsets. It is a non-invasive method that aids both adults and children in improving their brain function over time. When applied to the incorrect brain regions or when inappropriate neurofeedback protocols are employed for the subject's condition, traditional neurofeedback training may have unexpected effects. The 2-minute resting state eyes-open QEEG data is examined by Auto Train Brain's AI-assisted algorithms before each neurofeedback

session, and based on the results, the best neurofeedback protocol is provided for the following session. 1700 healthy individuals between 4- 80 years old provided Auto Train Brain with their norm data.

During the neurofeedback session, this norm data is utilized to compare the subject's QEEG values to those of normal persons. The most abnormal brain areas are subjected to neurofeedback training. Auto Train Brain was created with the goal of reducing negative effects and has been shown to be beneficial in earlier research and clinical trials for children (Eroğlu et al., 2021). It reads and interprets QEEG signals from 14 channels and delivers online, real-time visual and aural neurofeedback. A system and method for boosting learning capacity are offered inside this software program. Dyslexia is characterized by marked hypoactivation within the reading network, disrupted functional connectivity, and differences in structural connectivity in certain fiber tracts, thus we thought the following approach might be useful. This neurofeedback approach was previously tested on dyslexic children and patented (Eroğlu et al., 2021).

- Reduce absolute theta waves at FC5 if they are higher than the age-matched norm theta; and/or,
- Reduce absolute theta waves at T7- P7- O1 if they are higher than the age-matched norm theta; and/or,
- Find the channels with the highest absolute power of theta waves in the left hemisphere and reduce absolute theta; and/or,
- Find the channels with the highest absolute power of theta waves in the right hemisphere and reduce absolute theta.

A green arrow on the screen indicates a good reward, whereas a red arrow and a "beep" sound indicate negative feedback. The score displayed on the screen rises when a favorable reward is received. A red arrow is displayed on the screen if the person's slow brain waves are over the norm absolute theta, and the participant is instructed to try to transform it into a green arrow. A 30-minute neurofeedback session is applied first. A phoneme-grapheme matching alphabet teaching method is offered after the neurofeedback session that consists of visual and auditory cues. Applying multi-sensory learning right after neurofeedback makes semantic learning permanent as the brain becomes ready to learn new information due to neurofeedback. One of the key distinctions between the Auto Train Brain and other neurofeedback systems is that it integrates neurofeedback with multi-sensory learning concepts. The 14-channel neurofeedback methodology is also unique. As described by Eroğlu et al. (2021), lowering theta using Auto Train Brain reduces the disconnection syndrome in dyslexia. Theoretically, Auto Train Brain improves functional connectivity by lowering theta in the temporal region and/or the left hemisphere. Coherence across various brain areas is practically enhanced by lowering QEEG theta band power, as evidenced by the TILLS test findings.

2.4. Study Design

The Auto Train Brain solution is a mobile app that may be used at home. Each participant sat on a chair while the electrodes were inserted for the QEEG assessment. The distance between the participant and the mobile phone screen was 0.5 meters. The QEEG measurements were taken for 2 minutes before each neurofeedback session using the EMOTIV EPOC+ or EPOC-X and the Auto Train Brain app. The participant was requested to complete the QEEG measurement while relaxing and with their eyes open. This study has a limited sample size and is set up as a repeated measurement.

The survey to assess the app's feasibility and acceptance by the users consist of 6 questions: (1) Is the diagnosis shown on the mobile app correspond to that of the psychiatrists when they started using the neurofeedback application

(later on the diagnosis may change due to neurofeedback)? (2) Is it easier to use the app on a child? (3) Do you suggest this app to any other people? (4) Would you prefer to use this app for dyslexia classification at home or at school? (5) Do you think the price of acquiring the solution (EEG headset and software subscription) is high/ moderate/ low for neurofeedback and dyslexia/normal classification purposes? (6) Do you think the price of 1-time measurement (dyslexia/normal classification) at school is high/ moderate /low? 41 people participated in the survey. These are middle-aged parents (35- 45 years old, 60% female, 40% male). The socioeconomic situation of the parents was middle class. The parents have used the Auto Train Brain solution at home beforehand.

2.5. *The Test of Integrated Language & Literacy Skills (TILLS)*

The TILLS is a test for the assessment of oral and written language abilities in students 6–18 years of age. Published in 2016 (Nelson, Plante, Helm-Estabrooks, & Hotz, 2016), it is unique in the way that it is aimed to thoroughly assess skills such as reading fluency, reading comprehension, phonological awareness, spelling, as well as writing in school-age children. The test is originally developed in English. Turkish Dyslexia Association has translated and adapted it to Turkish. This test has been used for diagnosing learning disabilities. For 6-7 years old children, a TILLS descriptive score of less than 24 indicates a learning disability with 84% sensitivity and 84% specificity. For 8-11 years old children, a TILLS descriptive score of less than 34 indicates a learning disability with 88% sensitivity and 85% specificity.

The TILLS test has 2 dimensions (language and modality). For listening modality, it has (1) Vocabulary awareness, (2) Phonemic awareness, (6) Listening comprehension, (8) Following directions; for speaking modality, it has (4) Nonword repetition, (3) Story retelling, (13) Social communication; for reading modality, it has (10) Nonword reading, (11) Reading fluency, (7) Reading comprehension; for writing modality, it has (5) Nonword spelling, (12a) Written expressions- Word score, (12b) Written expression -discourse score, (12c) Written Expression – sentence combining score; for Memory, (14) Digit span forward, (15) Digit span Backward, (9) Delayed story retelling subtests. The TILLS descriptive point is the sum of all subtests' scores.

2.6. *Statistical Analysis*

Python/Google Collab, Sci-kit Learn, and TensorFlow ML libraries were used to conduct the statistical and data analysis. K-Folding, Cross-validation, and confusion matrix generator functions derived from ML libraries. Mat plot library stands for plotting learning, validation, and ROC curves. The averaged 2-minute QEEG band power data (continuous data) are acquired from fourteen electrode channels each session. For the whole data set, the Z-scores are computed using the equation $z = (x-m)/s$ for each QEEG band power (including the experiment and the control group) for all the data gathered from dyslexic and healthy children. Because EMOTIV does not offer Z-scores, m and s stand for the sample's mean and sample standard deviation, respectively. Outliers (>5 or -5) were removed from the analysis. The missing values were replaced with the mean of the featured data. The data is labeled by a computer scientist following the psychiatrist's diagnosis of the participant. The diagnosis of the participant did not change across neurofeedback sessions

although there may be some improvements in the brain signals. The dyslexic and healthy participants' data were balanced with pre-processing operations and adjustments. The binary classification with a supervised ML model is applied. The model output is the dyslexia positivity probability score. The ML model architecture is Artificial Neural Network. The model features made in the study are epoch 60, batch size 32, and loss as binary cross-entropy. The best model is selected among many other varying hidden layers and activation functions. The k-fold cross-validation technique has been used to evaluate the model with ten-folded cross-validation. This method is generally used to test model performance to estimate how well the model performs on unseen data. The overfitting is prevented by applying the dropout between layers. The results have been tested with an external test set which contains a diverse set of input data. The model is then converted to a TFLITE model and embedded into Auto Train Brain neurofeedback mobile app.

3. Results

In this research, we have designed a Machine Learning algorithm to classify dyslexia and assessed the effectiveness with a survey on dyslexic families. This study is a repeated assessment with a limited sample size (12,420 sessions) (207 participants: 96 of them have dyslexia, and 111 of them are typically developing).

The artificial neural networks achieved high accuracy and low-loss functions. The results show the performance of suggested preprocessing methods and the artificial neural network models achieved 98.8% accuracy with 0.05 loss with a 95% of confidence interval (CI, k-fold cross-validation applied), which is the state-of-the-art in literature for dyslexia biomarker detection with EEG scans. In addition to that, this study concludes models which have additional preprocessing techniques like minimum-maximum scaler, standard scaler, etc. reduced the accuracy slightly from 98.8% to 98.63%. F1 score and loss become 0.986 and 0.07 respectively.

Table 1. Top 3 ANN results on the data set to detect the dyslexia biomarkers

Model Architecture	Accuracy	F1	Loss	AUC
1 st tanh , 2 nd softsign, 3 rd tanh , 4 th sigmoid,	0.9880	0.988	0.05	0.98
min max Scaler + 1 st tanh , 2 nd softsign 3 rd tanh , 4 th sigmoid	0.9863	0.986	0.07	0.96
1 st tanh , 2 nd softsign 3 rd linear , 4 th sigmoid	0.9818	0.982	0.09	0.95

In Table 1, the model architecture describes the activation functions at each level of ANN. Even though this study achieves high-accuracy results, it is also important to examine the cases of people who were misclassified by the model. According to the confusion matrix in Figure 1, the model claims 72 non-dyslexic people's sessions as dyslexic sessions. On the other hand, the model states that 27 dyslexic people's sessions as non-dyslexic sessions out of 8301 samples. In this study, the number of false negatives is more important than the number of false positives. Although, misclassifying

non-dyslexic people sessions as dyslexic -false positives- sessions would not create problems when it comes to the neurofeedback system of Auto Brain Train, but it is essential to consider dyslexic people's sessions classified as non-dyslexic -false negative- session predictions.




Figure 1. Confusion matrix result of the best model

The receiver Operating Characteristic (ROC) Curve, which stands for sensitivity and specificity, of the ANN model indicates the capability of the model to discriminate dyslexia and the results were very promising (Figure 2). This study achieved a 0.98 area under curve (AUC) score. According to the ROC Curve definition, an AUC value between 0.8 and 0.9 is considered excellent, and more than 0.9 is outstanding.




Figure 2. Receiver Operating Characteristic Curve

The survey results (N= 41) indicate that 96% of the respondents found the results of the diagnosis correspond to that of the psychiatrists; 92% of the respondents find the app easily usable at home; 82.9% of the respondents suggest the other users use the app. 50% of the respondents prefer to use the app for dyslexia/normal classification at home. 75% of the respondents find the total cost of ownership (the price of the EEG headset and the software subscription) high for usage at home. 78% of the respondents think the price of the one-time measurement (dyslexia/normal classification) at school is not expensive (Table 2, figure 3). It should be noted that this survey was completed by the already existing users/families of Auto Train Brain who were already diagnosed with dyslexia.

Table 2. Survey results on the usage of dyslexia diagnosis mobile app

Survey Questions	Results
Feasibility	96%
Acceptability	92%
Referral	83%
Prefer dyslexia detection usage at home	50%
Price high for neurofeedback at home	75%
Price not high for one-time measurement at school	78%




Figure 3. Survey results as a bar chart




Figure 4. Dyslexic data correlation matrix of features *a*(Heatmap), *b*(Filtered Heatmap)




Figure 5. Non – Dyslexic data correlation matrix of features *a*(Heatmap), *b*(Filtered Heatmap)

4. Discussion

The novelty of this research is that Z-score normalization of 14- channel QEEG data has led the high-accuracy results and the ML model is embedded into Auto Train Brain mobile app for everyday usage. The feasibility, acceptability, and economic impact of the mobile app on dyslexic families are assessed with a survey in the real life afterward.

The QEEG scan correlation between dyslexic and non-dyslexic data is clearly shows that dyslexic data has higher correlation between Theta, Beta and Gama signals than the non-dyslexic data. According to our measurement and calculations, neuron weight is quite high on the features which are effective to classification are T_T7, T_P8, T_P7, and T_O2.

This is the first research in the literature where ANN yields high accuracy results (98.88%) with z-scored QEEG data for dyslexia biomarker detection. As we collect QEEG data from 14 channels, it is expected that data may be correlated, and the covariance matrix is not diagonal. ANN still performs well under these conditions with QEEG data. We have standardized the data with Z-score calculation and eliminated the outliers. We have also balanced the data sets of the experimental and the control group. Eliminating the noise and outliers in the data helped to achieve such high accuracy.

On EEG scan data sets, Al-Barhamtoshy and Motaweh (2017) used ANN (89.6% percent accuracy). Karim et al. (2013) employed a Multilayer Perceptron to detect dyslexia signs by collecting brain waves in the resting state (85% accuracy). Using machine learning, Frid and Breznitz (2018) investigated the variations in ERP signals between dyslexic and proficient readers (with ANN, 78% accuracy). O. L. Usman et al. (2020) achieved state-of-the-art CNN architecture with MRI scans at 84.6%.

Our research with QEEG data with ANN yielded 98.8% accuracy which is the highest accuracy in the literature with K-fold cross-validation. Even better results could be achieved by using different types of ANNS such as generative-adversarial networks, convolutional neural networks, and recurrent neural networks.

The machine learning model created with ANN is converted to a TFLITE model and embedded into the Android Mobile Application of Auto Train Brain on Android and iOS (Figure 5). In this way, the diagnosis of dyslexia with a high accuracy rate becomes possible with 2-minute resting-state QEEG data collected from Auto Train Brain (Figure 6, figure 7).



Figure 5. A dyslexic child who uses Auto Train Brain mHealth app




Figure 6. The cognitive reports automatically generated after 2-minute QEEG Measurement




Figure 7. The example cognitive reports that have been generated after a 2-minute QEEG Measurement

With the advancement of technologies, dyslexia detection starts at the age of 4 with 2-minute resting-state QEEG data with Auto Train Brain. It does not take much amount of time and money for the family to get this assessment. This measurement is standardized and an objective assessment, and not tiring for a child and it is definitely not as scary as visiting a psychiatrist. The child sits quietly for 2 minutes in a chair, the EEG headset is placed in 2 minutes, and the resting state EEG measurement was accomplished in 2 minutes, the results (whether he has dyslexia or not) are

immediately ready and can be seen on the screen. If there would be a possibility of dyslexia, remedial actions can be taken as soon as possible. The families found the mobile app a feasible and acceptable solution for dyslexia detection. If children with dyslexia will continue their dyslexia training with Auto Train Brain at home, the diagnosis of dyslexia becomes very economic for the families. If it is a first-time measurement, they can get an assessment with Auto Train Brain at school which will cost less than buying the EEG equipment and the software subscription. In this way, the first-time assessments (before neurofeedback) become very economic for the families too. The respondents in the survey were current users of Auto Train Brain, and they prefer to use technological solutions for improving the symptoms of dyslexia. The awareness of dyslexia and the possible training methods are not high in society. There are people who do not trust technological solutions and want to continue with traditional training methods for their dyslexic children. Although technology is ready to diagnose dyslexia with a high accuracy rate and improve the symptoms of dyslexia with mobile health solutions, families still think the economic impact of the whole solution is high to use at home. Learning disabilities in generations impact the SES of families. The families who participated in our survey have middle SES, they still find the solution expensive although they have bought it. Either EEG headset prices should come down to make it possible for all dyslexic families to use them at home, or the solution should also be available at school. As the dyslexia market is a niche (only 10% of society has dyslexia), EEG headset prices may not possibly be down in the short run. After using the software for a longer period, families request other objective assessments to see the improvements after neurofeedback training. So, dyslexia biomarker detection becomes not enough for them, families require a training roadmap for their child to overcome this condition as fast as possible. Using ML methods to predict the next possible training or the level/scale of dyslexia will be our next challenge to add to the mobile app.

4.1. Limitations of the Study

The study's first drawback is that it only includes 207 participants and 12,420 times the same individuals provided measurements. The trial should have included more participants; that would have been optimal. The likelihood of placebo effects is the study's second restriction. Children who get one-on-one interactions and specialized therapies may enhance their functioning primarily due to the social and environmental influence of those interventions, according to Gaab et al. (2019). Because the control group was not given an alternative intervention, placebo effects might be a substantial source of improvement.

The study's third drawback is the maturation effects. Throughout their growth, all children's brains undergo major changes. As a result, QEEG modifications are anticipated to be influenced by maturation during the next six months.

5. Conclusions

Machine learning methods could provide new insights into the treatment of dyslexia. In this work, we created a classification scheme for dyslexia diagnosis based on generated Z-scores from QEEG data and ANN. The results imply that the rehabilitation process will undergo a paradigm change.

Families will gain quicker, objective evaluation of their dyslexic children by using Auto Train Brain at home or at school as finding an expert in dyslexia may not be easy and reachable for families.

Ethics approval and consent to participate

After the experimental procedure was explained to them by the research ethics committee, the study protocol was approved by the Yeditepe University Ethics Committee, and the clinical trial was registered with the Turkey Pharmaceuticals and Medical Devices Agency, all of the participants gave their informed consent (Nbr: 71146310-511.06,2.11.2018).

Availability of data and material

The datasets generated during and/or analyzed during the current study are available from the corresponding author on reasonable request.

Code availability

None.

Acknowledgments

We want to express our sincere appreciation to the families that took part in the study; without their commitment and help, we might not have been able to finish it.

Declarations

Competing interests

None

Funding

None

Authors' contributions

G.E. wrote the main manuscript text including the title, abstract, introduction, materials and methods, results, and discussion. G.E prepared tables and figures.

M.K. wrote the results and enhanced the materials and methods with the ANN method and algorithm. M.K. prepared tables and figures.

B.K. collected the data from normal children.

All authors reviewed the manuscript.

References

- Al-Barhamtoshy, H. M., & Motaweh, D. M. (2017, February). Diagnosis of Dyslexia using computation analysis. In 2017 International Conference on Informatics, Health & Technology (ICIHT) (pp. 1-7). IEEE.
- American Psychiatric Association. (2013). Anxiety disorders. In Diagnostic and statistical manual of mental disorders (5th ed.). <https://doi.org/10.1176/appi.books.9780896425596.dsm05>
- Badcock, N. A., Mousikou, P., Mahajan, Y., De Lissa, P., Thie, J., & McArthur, G. (2013). Validation of the Emotiv EPOC® EEG gaming system for measuring research quality auditory ERPs. *PeerJ*, 1, e38.
- Benfatto, M., Öqvist Seimyr, G., Ygge, J., Pansell, T., Rydberg, A., & Jacobson, C. (2016). Screening for dyslexia using eye tracking during reading. *PloS one*, 11(12), e0165508.
- Bosch-Bayard, J., Peluso, V., Galan, L., Valdes Sosa, P., & Chiarenza, G. A. (2018). Clinical and electrophysiological differences between subjects with dysphonetic dyslexia and non-specific reading delay. *Brain sciences*, 8(9), 172.
- Chabot, R. J., di Michele, F., Prichep, L., & John, E. R. (2001). The clinical role of computerized EEG in the evaluation and treatment of learning and attention disorders in children and adolescents. *The Journal of neuropsychiatry and clinical neurosciences*, 13(2), 171-186.
- Chen, H., Uddin, L. Q., Guo, X., Wang, J., Wang, R., Wang, X., ... & Chen, H. (2019). Parsing brain structural heterogeneity in males with autism spectrum disorder reveals distinct clinical subtypes. *Human brain mapping*, 40(2), 628-637.
- García Chimeno, Y., Garcia Zapirain, B., Saralegui Prieto, I., & Fernandez-Ruanova, B. (2014). Automatic classification of dyslexic children by applying machine learning to fMRI images. *Bio-medical materials and engineering*, 24(6), 2995-3002.
- Coben, R., Wright, E. K., Decker, S. L., & Morgan, T. (2015). The impact of coherence neurofeedback on reading delays in learning disabled children: A randomized controlled study. *NeuroRegulation*, 2(4), 168-168.
- Dehaene, S. (2005). Evolution of human cortical circuits for reading and arithmetic: The “neuronal recycling” hypothesis. *From monkey brain to human brain*, 133-157.
- Dehdar, K., Mahdidoust, S., Salimi, M., Gholami-Mahtaj, L., Nazari, M., Mohammadi, S., ... & Raoufy, M. R. (2019). Allergen-induced anxiety-like behavior is associated with disruption of the medial prefrontal cortex-amygdala circuit. *Scientific reports*, 9(1), 1-13.
- Dunn, L. M., & Dunn, L. M. (1965). Peabody picture vocabulary test--.
- Eroğlu, G., Gürkan, M., Teber, S., Ertürk, K., Kırmızı, M., Ekici, B., ... & Çetin, M. (2020). Changes in EEG complexity with neurofeedback and multi-sensory learning in children with dyslexia: A multiscale entropy analysis. *Applied Neuropsychology: Child*, 1-12.
- Eroğlu, G., Teber, S., Ertürk, K., Kırmızı, M., Ekici, B., Arman, F., ... & Çetin, M. (2021). A mobile app that uses neurofeedback and multi-sensory learning methods improves reading abilities in dyslexia: A pilot study. *Applied*

Neuropsychology: Child, 1-11.

- Facoetti, A., Lorusso, M. L., Paganoni, P., Cattaneo, C., Galli, R., Umiltà, C., & Mascetti, G. G. (2003). Auditory and visual automatic attention deficits in developmental dyslexia. *Cognitive brain research*, 16(2), 185-191.
- Frid, A., & Manevitz, L. M. (2018). Features and machine learning for correlating and classifying between brain areas and dyslexia. *arXiv preprint arXiv:1812.10622*.
- Forgy, E. W. (1965). Cluster analysis of multivariate data: efficiency versus interpretability of classifications. *biometrics*, 21, 768-769.
- Gaab, J., Kossowsky, J., Ehlert, U., & Locher, C. "Effects and components of placebos with a psychological treatment rationale—three randomized-controlled studies". *Scientific reports*, 9(1) ((2019): 1-8.
- Giofrè, D., Toffalini, E., Provazza, S., Calcagnì, A., Altoè, G., & Roberts, D. J. (2019). Are children with developmental dyslexia all the same? A cluster analysis with more than 300 cases. *Dyslexia*, 25(3), 284-295.
- Harrar, V., Tammam, J., Pérez-Bellido, A., Pitt, A., Stein, J., & Spence, C. (2014). Multi-sensory integration and attention in developmental dyslexia. *Current Biology*, 24(5), 531-535.
- Handler, S. M., Fierson, W. M., & Section on Ophthalmology and Council on Children with Disabilities, American Academy of Ophthalmology, American Association for Pediatric Ophthalmology and Strabismus, and American Association of Certified Orthoptists. (2011). Learning disabilities, dyslexia, and vision. *Pediatrics*, 127(3), e818-e856.
- Heim, S., Tschierse, J., Amunts, K., Wilms, M., Vossel, S., Willmes, K., Grabowska, A., & Huber, W. (2008). Cognitive subtypes of dyslexia. *Acta neurobiologiae experimentalis*, 68(1), 73–82.
- Jastak, S., & Wilkinson, G. S. (1984). The wide range achievement test-revised. Jastak Associates.
- Khan, R. U., Cheng, J. L. A., & Bee, O. Y. (2018). Machine learning and Dyslexia: Diagnostic and classification system (DCS) for kids with learning disabilities. *International Journal of Engineering & Technology*, 7(3.18), 97-100.
- Karim, I., Abdul, W., & Kamaruddin, N. (2013, March). Classification of dyslexic and normal children during resting condition using KDE and MLP. In 2013 5th International Conference on Information and Communication Technology for the Muslim World (ICT4M) (pp. 1-5). IEEE.
- Kaufman, A. S. (1994). Intelligent testing with the WISC-III. John Wiley & Sons.
- King, W. M., Giess, S. A., & Lombardino, L. J. (2007). Subtyping of children with developmental dyslexia via bootstrap aggregated clustering and the gap statistic: comparison with the double-deficit hypothesis. *International journal of language & communication disorders*, 42(1), 77-95.
- Klimesch, W., Doppelmayr, M., Wimmer, H., Gruber, W., Röhm, D., Schwaiger, J., & Hutzler, F. (2001). Alpha and beta band power changes in normal and dyslexic children. *Clinical Neurophysiology*, 112(7), 1186-1195.
- Koenig, O., Reiss, L. P., & Kosslyn, S. M. (1990). The development of spatial relation representations: Evidence from studies of cerebral lateralization. *Journal of Experimental Child Psychology*, 50(1), 119-130.
- Kuhl, U., Neef, N. E., Kraft, I., Schaadt, G., Dörr, L., Brauer, J., ... & Skeide, M. A. (2020). The emergence of dyslexia in the developing brain. *Neuroimage*, 211, 116633.
- Lallier, M., Tainturier, M. J., Dering, B., Donnadieu, S., Valdois, S., & Thierry, G. (2010). Behavioral and ERP evidence for amodal sluggish attentional shifting in developmental dyslexia. *Neuropsychologia*, 48(14), 4125-4135.
- Lum, J. A., Ullman, M. T., & Conti-Ramsden, G. (2013). Procedural learning is impaired in dyslexia: Evidence from a

meta-analysis of serial reaction time studies. *Research in developmental disabilities*, 34(10), 3460-3476.

- Lyon, G. R., Shaywitz, S. E., & Shaywitz, B. A. (2003). A definition of dyslexia. *Annals of dyslexia*, 53(1), 1-14.
- Paulesu, E., Frith, U., Snowling, M., Gallagher, A., Morton, J., Frackowiak, R. S., & Frith, C. D. (1996). Is developmental dyslexia a disconnection syndrome? Evidence from PET scanning. *Brain*, 119(1), 143-157.
- Rippon, G., & Brunswick, N. (2000). Trait and state EEG indices of information processing in developmental dyslexia. *International Journal of Psychophysiology*, 36(3), 251-265.
- Schrank, F. A. (2011). *Woodcock-Johnson III Tests of Cognitive Abilities*.
- Snowling, M. J., & Hulme, C. (2012). Annual Research Review: The nature and classification of reading disorders—a commentary on proposals for DSM-5. *Journal of child psychology and psychiatry*, 53(5), 593-607.
- Spoon, K., Crandall, D., & Siek, K. (2019, June). Towards Detecting Dyslexia in children's handwriting using neural networks. In *Proceedings of the International Conference on Machine Learning AI for Social Good Workshop*, Long Beach, CA, USA (pp. 1-5).
- Stein, J., & Walsh, V. (1997). To see but not to read; the magnocellular theory of dyslexia. *Trends in neurosciences*, 20(4), 147-152.
- Sheremet, A., Zhou, Y., Kennedy, J. P., Qin, Y., Burke, S. N., & Maurer, A. P. (2018). Theta-gamma coupling: a nonlinear dynamical model. *BioRxiv*, 304238.
- Tamboer, P., et al. "Machine Learning and Dyslexia: Classification of Individual Structural Neuro-Imaging Scans of Students with and without Dyslexia." *NeuroImage: Clinical*, Elsevier, 29 Mar. 2016.
- Usman, O. L., & Muniyandi, R. C. (2020, May 20). *CryptoDL: Predicting dyslexia biomarkers from encrypted neuroimaging dataset using energy-efficient residue number system and deep convolutional neural network*. MDPI. Retrieved July 28, 2022.
- Nelson, N. W., Plante, E., Helm-Estabrooks, N., & Hotz, G. (2016). *Test of integrated language and literacy skills (TILLS)*. Baltimore, MD: Brookes.
- Wang, J., Barstein, J., Ethridge, L. E., Mosconi, M. W., Takarae, Y., & Sweeney, J. A. (2013). Resting state EEG abnormalities in autism spectrum disorders. *Journal of neurodevelopmental disorders*, 5(1), 1-14.
- Wagner, R. K., Torgesen, J. K., Rashotte, C. A., & Pearson, N. A. (1999). *Comprehensive test of phonological processing: CTOPP*. Austin, TX: Pro-ed.
- Wang, J., Xing, G. G., Li, X., & Wan, Y. (2016). Enhanced gamma oscillatory activity in rats with chronic inflammatory pain. *Frontiers in neuroscience*, 10, 489.
- Wechsler, D. (1999). *Wechsler abbreviated scale of intelligence--*.