

[Open Peer Review on Qeios](#)

# Testing of Device from Other Lot/Batch Retained by Manufacturer

National Cancer Institute

## Source

National Cancer Institute. *Testing of Device from Other Lot/Batch Retained by Manufacturer*. NCI Thesaurus. Code C139448.

The investigation employed relevant empirical testing of the device of another lot or batch than that of the suspected device in the reported adverse event in order to support the identification of possible causes for the adverse event. This includes devices without a lot/batch designation. Testing was performed using the device retained by the manufacturer (i.e. was not shipped). Relevant testing would typically be based on test methods used for evaluating safety and performance as described in the latest relevant standards.